

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
DataMaster cdm S/N <u>130183</u>	
Name <u>HUTT, William T.</u> ID# <u>3359</u> Date <u>05/03/10</u>	
A	Agency <u>Homer Police Dept./ Jail</u> Phone # <u>235-3150</u>
Instrument Location <u>4060 Heath St. Homer, AK</u> <u>235-3151 (fax)</u>	
<u>99603</u>	
Alco S/N <u>A090740</u> Target Value <u>0.084</u> High Pressure <u>900 psi</u>	
B	Alco Test Values <u>0.082</u> <u>0.084</u>
	1 <sup>st</sup> Alco 2 <sup>nd</sup> Alco
Signature <u>William T. Hutt</u> <u>CB</u>	
<u>5/24/10</u>	
(OVER)	

I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (i) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 10th day of June, 2010.

Carolyn M. Noland (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130183

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130183

MAY 03, 2010

OPERATOR'S NAME:

HUTT, WILLIAM T

OPERATOR'S NUMBER: 3359

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

ETCH

O.L. #: 01234567

DEPT/AGENCY: HOM1

CASE/REPORT: 10-0000

TEST TYPE: V

ALCO TARGET VALUE: 0.004

ALCO S/N: A090740

### — BREATH ANALYSIS —

.004 ADJUSTED FOR 30.25 in

ALCO TARGET .004 18:55

BLANK TEST .000 18:56

INTERNAL STANDARD .000 VERIFIED 18:56

ALCO TV 30.25 in .002 18:56

BLANK TEST .000 18:57

SUBJECT SAMPLE .000 18:58

BLANK TEST .000 18:59

ALCO TV 30.25 in .004 19:00

BLANK TEST .000 19:00

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130183

MAY 03, 2010

TIME 18:51

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 30.24 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmnop  
qrstuvwxyz{|}~